ANALYSIS OF TEMPERATURE CHARACTERISTICS OF THE PLUNGER PAIR OF THE CERAMIC PLUNGER PUMP FOR HEMODIALYSIS

by

Song LIU^{*a,b,c*}, Ling PENG^{*b*}, Xian HUANG^{*a**}, Bing Rong LIU^{*b*}, Meng Jie TU^{*c*}, and Yi Han WANG^{*c*}

 ^a School of Precision Instrument and Opto-Electronics Engineering, Tianjin University, Tianjin, China
 ^b Jiangxi Sanxin Medical Technology Co., Ltd., Nanchang, China
 ^c Jiangxi Shengdankang Medical Technology Co., Ltd., Nanchang, China

> Original scientific paper https://doi.org/10.2298/TSCI231213048L

Through the design of the ceramic plunger pump for hemodialysis, the mathematical model of the water film of the plunger pair of the ceramic plunger pump was built. The distribution characteristics of the speed and pressure of the water film of the plunger pair were analyzed, and the temperature distribution characteristics of the water film surface were obtained. The effects of inlet pressure, inlet temperature, rotational speed, and chamber temperature on the water film temperature of the plunger pair were studied. Experiments and analysis show that the closer the temperature of the plunger pair is to the inlet of purified water, the smaller the temperature change. The farther the ceramic plunger rod is from the inlet of purified water, the greater the temperature change, and the main influencing points are concentrated inside the plunger pair. Therefore, maintaining a certain amount of purified water temperature, pressure, and flow rate for lubrication is crucial for the service life of ceramic piston pumps.

Key words: hemodialysis, ceramic plunger pump, plunger pair, water film, temperature characteristic

Introduction

Blood purification is one of the conventional treatment methods for chronic renal failure, and in the course of blood purification treatment, the usual treatment modes are hemodialysis and peritoneal dialysis. In the mode of hemodialysis treatment, treatment equipment such as hemodialysis machine will be used, because hemodialysis requires the mixing ratio of hemodialysis concentrate to enter the dialyzer with a certain flow rate and pressure for dispersion, convection, ultrafiltration and other functions [1]. In use, a certain volume of saturated concentrate A and saturated concentrate B is quantitatively extracted by a ceramic plunger pump by matching the concentrate solution of hemodialysis. At the same time, purified water is entered into the lubrication chamber of the ceramic plunger pump, the plunger pair is an important moving part of the ceramic plunger pump, which directly affects the temperature characteristics of the plunger pair movement, resulting in aging and wear of the plunger rod, *etc.*, which will affect

^{*}Corresponding author, e-mail: huangxian@tju.edu.cn

the pressure distribution and flow of the plunger pump. The high temperature will also cause the thermal deformation of the lubricating auxiliary surface for a long time, which will affect the assembly gap and cause the failure of the surface material of the moving auxiliary, resulting in the normal use of the ceramic plunger pump, so it is necessary to analyze and calculate [3].

In recent years, with the gradual development and application of polymer materials and engineering ceramic materials, various universities and research institutions have conducted development and research on ceramic plunger pumps, such as coke Juan, etc. to analyze the cavitation phenomenon and friction loss temperature of pure water pressure plunger pumps, and to study the friction characteristics of ceramic materials, stainless steel and modified polyether ether ketone under water lubrication. Moreover, relevant tests were carried out on its dynamic characteristics through design experiments Hu and Fei [4]. Xiong et al. [5] optimized the design and strength analysis of the key structural components of the hydraulic piston pump by using basic theoretical design and combined with simulation analysis. By establishing the actual output flow model of the hydraulic piston pump, Yin and Nie [6] optimized the structure parameters of the valve plate under the condition of considering the leakage flow of key friction pairs and the pre-lift pressure angle, and conducted numerical analysis on the motion state and position of the pump to study the reasons for the flow pulse generated by the hydraulic axial piston pump. Zhou and Bian [7] designed the aseptic precision filling zirconia ceramic valveless pump, established the relationship model between pressure and flow, and proposed the realization method of flow control and precision fine-tuning of valveless rotary pump. In summary, most researchers mainly establish a simplified motion model to optimize the design of friction characteristics and flow rate, and rarely analyze the influence of structural parameters of ceramic piston pumps on the temperature of moving pairs.

Combined with the characteristics of the analysis, in order to accurately describe the distribution characteristics and rules of the overall temperature of the water film lubrication of the plunger pair, the temperature characteristics of the purified water were fully considered in the way of combining the principles of thermodynamics and fluid mechanics, and the water film temperature and distribution rules along the axial direction of the plunger pump were analyzed and studied, especially the influence of the relevant structural parameters of the ceramic plunger pump on the entire temperature distribution [8].

System description

Design of ceramic plunger pump for hemodialysis

The plunger pump mainly consists of a motor, a motor fixing bracket, a rotating shaft sleeve, a pump head connecting rod, a ceramic plunger, a cylinder, a dialysate chamber, and a purified water chamber. The motor is mounted on a fixed bracket and rotates by driving a rotating shaft sleeve [9]. The rotating shaft sleeve is connected with the link rod of the pump head, and the ceramic plunger rotates in the cylinder barrel by rotating in the angle support bearing: a certain angle A is formed between the central axis of the ceramic plunger and the horizontal line, so that the cylinder of the ceramic plunger moves back and forth in a straight line, and a certain volume of chamber is formed inside to realize the night discharge and suction in the ceramic plunger pump, so as to achieve the purpose of the cannery. In the process of movement, the pump head is connected with the rod part and rotates in the form of an arc [10].

In the dialysate chamber of the hemodialysis ceramic plunger pump, the ceramic plunger rod cuts the volume of length, L, and depth, H, to form the liquid storage space of the plunger pump, that is, the liquid storage chamber. When the ceramic piston pump rotates in the cylinder, the liquid is in the state to be injected when the plane of the liquid storage chamber

is parallel to the inlet chamber. When the ceramic plunger rod is rotated 90°, the liquid enters the liquid reservoir and fills. When the rotation continues to 90°, a certain internal pressure is formed in the liquid storage chamber to be discharged. When the final rotation is 90°, the liquid is discharged from the drain port. At the same time, the purified water cavity continues to have liquid-flowing from the inlet and the outlet, so that the plunger rod keeps moisture lubrication, because the gap between the plunger rod and the cylinder is co-ordinated, the precision of the co-ordination is very high, long-term movement makes the plunger rod heat, so it is very important to maintain the lubrication between the plunger rod and the cylinder [11]. The structure of the ceramic plunger pump and the movement of the plunger pair are shown in fig. 1.



Figure 1. Ceramic plunger pump structure and plunger pair movement

Displacement of plunger movement

According to the status and characteristics of the motion of the ceramic piston pump, when the angle support bearing of the ceramic piston pump rotates to the horizontal position of 0°, that is, point P, the plunger pair is in the shortest position point, the pump motion working state is in the pumping state, the rotating rod is represented by PC, and the corresponding plunger rod is represented by CO. When the angle support bearing of the ceramic plunger pump is rotated to the horizontal position of 180°, that is, P' point, the plunger pair is in the longest position point, the working state is in the suction state, the rotating rod is represented by P'A, and the corresponding plunger rod is represented by AD and PC = P'A = R through the calculation of rotation 0-360° displacement OD = 2R*tan(a), at the same time, it can be concluded that the greater the tilt angle (a) of the plunger rod, the greater the suction corresponding to each suction beat, the smaller the tilt angle (a), the smaller the displacement, the smaller the suction volume [12].

Reynolds equation of water film for plunger auxiliary inlet

According to Reynolds equation obtained from the water film of the plunger pair of the ceramic plunger pump [13], the analysis method of the water film thickness is established in the reference plane, z = 0 as shown in the fig. 2. At any point a(x, y, 0) is projected on the upper and lower surfaces of the plunger pairs, denoted by $a_c(x, y, z_c)$ and $a_p(x, y, z_p)$, respectively.

$$h(x, y) = z_{c}(x, y) - z_{p}(x, y)$$
(1)



Figure 2. Schematic diagram of the water film plane of the plunger auxiliary

The Naiver-Stokes formula based on viscous flow is expressed:

$$\nabla(\rho v v) + \rho \frac{\partial v}{\partial t} = \nabla \sigma - \nabla p + \rho f$$
⁽²⁾

where $v \text{ [ms^{-1}]}$ is the water velocity vector of the plunger pair, $\rho \text{ [kgm^{-3}]}$ – the water density, $\sigma \text{ [Pa]}$ – the stress tensor of water, p [Pa] – the pressure in the flow field, and $f \text{ [Nm^{-3}]}$ – the volume force of water.

Considering that purified water is presented as a Newtonian fluid, there is a certain relationship between the applied tensor flow rate and the velocity vector:

$$\sigma = \mu \nabla v \tag{3}$$

Since the length and width of the purified water film are greater than its own thickness during the sliding pair movement, there are corresponding assumptions for this, and the viscous internal force between the water films will be greater than the inertial force, so $\nabla(\rho vv)$ and volume force ρf can be ignored. It is assumed that the fluid state in the water film is steady flow, expressed as $\rho(\partial v/\partial t)$. The liquid pressure value of the water film does not change the direction of the thickness of the water film, so the liquid pressure of the water film is expressed as p = p(x, y). The gradient of the water film liquid velocity vector in the water film thickness is greater than the gradient value of the ink length and width direction, expressed as:

$$\frac{\partial u, v}{\partial z} \rangle \rangle \frac{\partial u, v}{\partial x}, \ \frac{\partial u, v}{\partial z} \rangle \rangle \frac{\partial u, v}{\partial y}$$

Based on the aforementioned hypothetical analysis, the Naiver-Stokes publicity can be simplified:

$$\nabla p = \nabla(\mu \nabla v) \tag{4}$$

Temperature analysis

When the purified water passes through the gap of the ceramic plunger pump, the water film is rubbed against the surface of the ceramic plunger, and the heat generated will cause the water temperature to rise. The fluid unit is selected as the analysis object, and the object flows through the gap of the ceramic piston pump to produce energy changes: first, the change of purified water pressure will cause the loss of water film energy [14] and secondly, the friction of the viscous liquid produces the loss of energy. The final energy change of the cavity and plunger and the surrounding environment. During the process, the change of purified water pressure, viscous friction loss and environmental change affect the energy change of the analysis object, and the change of purified water energy is mainly reflected in the change of water film temperature Δt . Therefore, a mathematical model of the secondary temperature field distribution of the ceramic piston pump can be established to analyze the temperature change during the moving process.

Pressure drop energy loss in liquid inlet chamber

The energy loss caused by pressure changes will be converted into heat energy. The pressure difference between the pressure, P, of the ceramic piston pump in normal operation and the research object P(z) is derived and analyzed [15]:

$$\Delta p = (p_1 - p_2)\frac{z}{L} \tag{5}$$

In unit time, the energy value of the resulting pressure drop loss:

$$E_p = qv\Delta p \tag{6}$$

Changes in liquid friction energy loss

During the flow of purified water in the gap of the plunger, the friction loss will be converted into the heat energy entering the research object. The water film thickness, *s*, is very small, which can be determined as the velocity gradient in the thickness direction of the water film is the velocity difference between the upper and lower contact surfaces of the water film. Therefore, combined with Newton's law of internal friction [16], the shear stress of the water film can be obtained:

$$t_z = \frac{\varepsilon v}{h} - \frac{\left(p_1 - p_2\right)s}{2L} \tag{7}$$

The energy lost by liquid friction of the research object per unit time:

$$E_f = t_z v_z \mathrm{d}x \tag{8}$$

where dx represents the differential length of the research object.

Change of heat conduction energy loss

The flow form of purified water in the plunger pair of the ceramic plunger pump can be expressed as the flow of fluid between two parallel wall surfaces. The forms of heat transfer and exchange between purified water and the environment include heat convection, heat conduction and heat radiation. However, the value of energy exchange generated by heat conduction and heat radiation is small and can be ignored [17]. In this paper, only the influence of heat convection is considered, and the energy loss generated by the research object due to heat conduction factors in unit time can be expressed:

$$E_{\omega} = \alpha (T - T_{w}) \mathrm{d}x \tag{9}$$

where T is the temperature of the fluid, T_w – the average surface the transfer pair, and α – the surface heat transfer coefficient.

Surface temperature change of water film

The energy loss caused by liquid friction enters the water film of the plunger, and the change of water film energy can be expressed:

$$E = E_f + E_p - E_\varphi \tag{10}$$

The energy change value of the water film can be expressed as a change in temperature:

$$T = t_0 + \frac{\Delta E}{\rho c q \nu} \tag{11}$$

where t_0 is the initial state temperature of purified water, ρ – the density of purified water, and c – the mass heat capacity of purified water.

In the analysis, ignoring the energy transfer between the liquid and assuming that the temperature change of each fluid unit is independent of each other, the water film of the ceramic plunger pump can be divided into a number of finite units, and the temperature field characteristic curve of the water film of the plunger can be obtained by analyzing it.



Figure 3. Flow chart of structure field solver

Plunger temperature field analysis

According to the loading and constraint of the ceramic plunger cavity and the data transfer of the temperature field of the fluid structure, the fluid-structure coupling of the water film of the plunger can be calculated and analyzed. The flow of the structural field solver for the calculation of cavity and plunger deformation is shown in fig. 3.

Study on thermal deformation of ceramic plunger and cavity

Due to the time-varying characteristics of the temperature distribution of the ceramic plunger and the cylinder cavity, especially

> electric nstant

9.7

when the plunger pump works steadily, the heat exchange of the plunger pump tends to be stable, and the temperature change of the ceramic plunger and the cylinder cavity is not obvious. Due to the environmental requirements of high temperature resistance and corrosion resistance, the ceramic plunger and the cylinder cavity are made of Al_2O_3 material. At the same time, the outer chamber is made of PVDF material and the cylinder chamber is installed inside. The thermal deformation of ceramic piston pump cylinder and plunger was studied by simulation software. The material properties of the ceramic plunger and cylinder cavity involved in the simulation are shown in tab. 1.

Table 1. Material property parameters								
	Mod	Materials	Modulus of elasticity [Nm ⁻²]	Poisson's ratio	Mass density [gcm ⁻³]	Thermal expansion system [1 · 10 ⁻⁶ °C]	Thermal conductivity [Wm ⁻¹ K ⁻¹]	Die
	Ceramic plunger/ cylinder	Al ₂ O ₃	3.5-4 · 10 ¹¹	0.23	3.89	7.9	30	

Table 1. Material property parameters

chamber



Figure 4. Mesh division and thermal deformation of cavity; (a) cavity meshing and (b) cavity thermal deformation

Set the initial temperature to 20 °C, and calculate the deformation of the cylinder cavity and plunger when heated to 45 °C, respectively. The grid and results of thermal expansion analysis of the cavity are shown in fig. 4. The two holes of the cavity are purified water inlet and outlet holes and dialysate inlet and outlet holes, respectively. In fig. 4(a), the cylinder cavity is divided into 22014 tetrahedral mesh units and 33259 nodes. Figure 4(b) indicates that the thermal deformation of the cylinder cavity is concentrated in the part close to the contact of the moving pair.

Figure 5 shows the grid and results of thermal expansion analysis of the plunger. Figure 5(a) shows that the plunger is divided into 5351 tetrahedral grid units and 8724 nodes. Figure 5(b) shows that the thermal deformation of the ceramic plunger is concentrated in the moving contact of the plunger drainage fluid. The missing cylinder in the figure is the volume of the plunger chamber. The bottom hole is the mounting hole for the plunger to connect the movable rod.

Experimental analysis

The water film temperature of ceramic piston pump auxiliary was studied and the whole parameter was analyzed [18]. Combined with the designed ceramic piston pump working



Figure 5. Grid division and thermal deformation of ceramic plunger; (a) ceramic plunger meshing and (b) ceramic plunger thermal deformation

parameters, the maximum continuous pressure \leq 4bar, the maximum intermittent pressure of 6 bar, the speed range: 0-1000 rpm, the pump displacement: 0-195 µL/rpm (adjustable), the plunger length 50 mm, the tilt angle 0-15° (adjustable), the water film thickness of 20 µm.

According to the environmental conditions, the indoor working temperature is 20 °C. Ignore the temperature change or loss of the silicone pipe connection, and the temperature at the entrance and outside is 20 °C. According to the motion characteristics of the ceramic piston pump, the position of the maximum speed of the piston is selected for research and analysis, and the pressure fluctuation and other changing factors of the transient ceramic piston pump are ignored in the process. When the plunger is squeezing the liquid-flow and the plunger is pressing the liquid area at the bottom, the deeper the push, the more significant the temperature change of the water film on the plunger side [19].

When the ceramic piston pump moves, the moving medium into which the liquid-flows is purified water. The viscosity of purified water is affected by temperature, pressure and other factors [20]. If it is assumed that purified water is a liquid that cannot be compressed, the viscosity of purified water is significantly affected by temperature, and the viscosity characteristic curve of purified water is shown:

$$\mu = u_0 \mathrm{e}^{-a\Delta T} \tag{12}$$

where *a* is viscosity temperature coefficient of purified water, u_0 – the purified hydrodynamic viscosity at temperature T_0 , and ΔT – the temperature increment:

$$\Delta T = T - T_0 \tag{13}$$

The viscosity, density and other related physical indexes of purified water are directly related to the ceramic piston pump [21]. The density changes with temperature as shown in the figure, the density of water itself is not affected by the change of environmental pressure, and when the temperature changes, the density has little influence. In this analysis, the influence of temperature on the density of purified water is ignored [22].

Temperature distribution of the fluid along the axial direction of the plunger

The fig. 6 shows a schematic diagram of the reciprocating working area of the ceramic piston pump. The purified water flows vertically down the plunger pair, and the bottom of the liquid inlet is selected as the co-ordinate origin, while co-ordinate x represents the position of the water film of the research object.





Figure 7 shows the axial temperature distribution of the water film of the auxiliary plunger. When the purified water enters from the inlet and moves, the water temperature keeps rising and reaches the end, it can reach 30 °C. The calorific value of purified water is related to the viscosity, and the viscosity decreases with the increase of temperature. The reduction of viscosity will also cause the corresponding shear stress to change, reducing friction, and the



Figure 7. Axial temperature distribution of water film in ceramic piston pump



Figure 8. Working test diagram

work generated heat is also reduced, resulting in little temperature change of the purified water. During operation, the inlet will continue to flow into new purified water, which is the heat that generated friction before and will be taken away by the new purified water, so the heat conduction on the surface of the cavity will also be reduced, thus reducing the temperature growth of the water film [23].

Analysis of influencing factors of water film temperature of plunger pair

The distribution of water film temperature has a direct influence on the working pressure, working speed and tilt angle of the ceramic piston pump. The design of ceramic piston pump is studied and analyzed, and the influence of single parameter change in different categories on the temperature distribution of water film is analyzed. Speed, working wa-

ter pressure and tilt angle are tested by self-made circuit boards, motor drivers and ceramic plunger pumps, as shown in fig. 8.



Influence of rotational speed on water film temperature

For the designed ceramic piston pump, the speed is set to 300 rpm, 500 rpm, and 800 rpm, respectively. The test time is 5 minutes, it can be seen that the higher the rotational speed, the higher the outlet temperature, as shown in fig. 9.

Influence of working pressure on water film temperature

The normal continuous working pressure of the ceramic plunger pump is 4 bar. When the pressure changes to 2 bar and 6 bar, the temperature distribution changes are obtained through analysis. The water temperature at the outlet of the plunger pair of the ceramic plunger pump increases with the increase of the pressure, and the temperature also decreases with the decrease of the pressure, as shown in fig. 10.

Influence of tilt angle on water film temperature

The tilt angle of the ceramic piston pump ranges from 0-15°. Through the aforementioned analysis, it is known that the greater the tilt angle, the greater the offset position angle, the greater the corresponding partial friction force, the higher the water temperature, the longer the travel of the movement, and the more volume of water pumped, as shown in fig. 11. However,



Figure 9. Temperature distribution of plunger pair at different speed states



Figure 10. Temperature distribution of plunger pair under different working pressures



of plunger pair at different tilt angles

the dip angle has little effect on the surface temperature of the water film and is not the main factor affecting the temperature distribution of the plunger.

Conclusion

The speed of the ceramic piston pump directly affects the temperature of the plunger pair, the faster the speed, the higher the water film temperature. Secondly, the higher the working pressure, the higher the water film temperature of the plunger pair, and the tilt angle of the plunger pump has no great influence on the temperature change of the plunger pair. The closer the temperature of the plunger pair is to the purified water inlet, the smaller the temperature change, the farther the distance between the ceramic plunger rod and the purified water inlet, the greater the temperature change, therefore, to maintain a certain amount of purified water temperature, pressure and flow lubrication, the service life of the ceramic plunger pump is very important.

Acknowledgment

Nanchang City science and technology major project Hongkezi [2023] No. 137.

References

- Liu, S., et al., Design and Experimental Study of Portable Extracorporeal Circulation Pipe-Line Performance Test System, Chinese Journal of Medical Instrumentation, 46 (2022), 02, pp. 164-167
- [2] Yin, Y., et al., Research on Temperature Characteristics of Piston Pair of PP Axial Piston Pump, China Mechanical Engineering, 26 (2015), 08, pp.1073-1077+1084
- [3] Zhang, T., Liu, X., Reliability Design for Impact Vibration of Hydraulic Pressure Pipe-Line Systems, Chinese Journal of Mechanical Engineering, 26 (2013), 5, pp, 1050-1055
- [4] Hu, Z., Fei, H., Thermal Coupling Effect Analysis of Geotechnical Engineering Reinforcement Materials, *Thermal Science*, 25 (2021), 6A, pp. 4067-4073
- [5] Xiong, D.-K., Improvement of Newton's Law of Internal Friction and Darcy's Law, Geotechnical Engineering Technology, 18 (2004), 06, pp. 275-278
- [6] Yin, F., Nie, S., Pre-Boost Angle of Valve Plate of Hydraulic Axial Piston Pump, Journal of Beijing University of Technology, 41 (2015), 09, pp. 1281-1288
- [7] Zhou, H., Bian, S., Design and Characteristic Analysis of Zirconia Ceramic Valveless Pump for Aseptic Precision Filling, *Packaging Engineering*, 40 (2019), 09, pp. 161-166
- [8] Li, H., Research on Visual Simulation and Application of Marine oil System, Ph. D. thesis, Dalian Maritime University, Dalian, China, 2020
- [9] Zhang, K., Design of Material Transportation System Based on Electromagnetic Module, Ph. D. thesis, Huazhong University of Science and Technology, Huazhong, China, 2015
- [10] Ameur, H., et al., Numerical Investigation of The Performance of Perforated Baffles in a Plate-Fin Heat Exchanger, *Thermal Science*, 25 (2021), 5B, pp. 3629-3641
- [11] Xiong, D., Improvement of Newton's Law of Internal Friction and Darcy's Law, Geotechnical Engineering Technology, 18 (2004), 06, pp. 275-278
- [12] Jiang, Q., et al., Numerical Analysis of the Water Film Characteristics in the Eccentric State of a Radial Piston Pump, *IEEE Access*, 6 (2018), 2, pp. 15274-15282
- [13] Ivantysyn, J., Ivantysynova, M., Hydrostatic Pumps and Motors, Academic Books International, New Delhi, India, 2001, pp.140-143
- [14] Xu, B., Investigation on the Radial Micromotion about Piston of Axial Piston Pump, Chin. J. Mech. Eng., 26 (2013), 2, pp. 325-333
- [15] Tanaka, K., et al., Characteristics on Sliding Surfaces between Piston and Cylinder in a Piston Pump and Motor: Effects of Running-in, Profile of Pistonp and Stiffness, JSME International Journal Series C-Mechanical Systems Machine Elements and Manufacturing, 64 (1998), 626, pp. 3959-3967
- [16] Haidak, G., Wang, D. E. S., Research on the Thermo-Elastic Deformation and Fracture Mechanism of the Slipper Retainer in the Axial Piston Pumps and Motors, *Engineering Failure Analysis*, 100 (2019), June, pp. 259-272
- [17] Wang, D. Y., et al., Research on the Fluid Film Lubrication between the Piston-Cylinder Interface, AIP Advances, 8 (2018), 10, 17
- [18] Petrovic, R., Mathematical Modelling and Experimental Research of Characteristic Parameters Hydrodynamic Processes of a Piston Axial Pump, *Strojniski Vestnik Journal of Mechanical Engineering*, 55 (2009), 4, pp. 224-229
- [19] Xu, B., et al., Simulation Research on Distribution Method of Axial Piston Pump Utilizing Pressure Equalization Mechanism, Proceedings of Institution of Mechanical Engineers Part C, Journal of Mechanical Engineering Science, 227 (2013), 3, pp. 459-469
- [20] Jiang, J., et al., The Impact of Bushing Thickness on the Piston/Cylinder Interface in Axial Piston Pump, IEEE Access, 7 (2019), Feb., pp. 24971-24977
- [21] Song, Y. H., et al., A Numerical Study on Influence of Temperature on Lubricant Film Characteristics of the Piston/Cylinder Interface in Axial Piston Pumps, *Energies*, 11 (2018), 7, pp, 121-132
- [22] Wang, D. Y., et al., Research on the Fluid Film Lubrication Between the Piston-Cylinder Interface, AIP Advances, 8 (2018), 10, pp. 1-16
- [23] Sun, L., Zhang, C. L., Evaluation of Elliptical Finned-Tube Heat Exchanger Performance Using CFD and Response Surface Methodology, *International Journal of Thermal Sciences*, 75 (2014), Jan., pp. 45-53

Paper submitted: December 13, 2023 Paper revised: December 21, 2023 Paper accepted: January 5, 2024 © 2024 Society of Thermal Engineers of Serbia Published by the Vinča Institute of Nuclear Sciences, Belgrade, Serbia. This is an open access article distributed under the CC BY-NC-ND 4.0 terms and conditions